

पिंपरी चिंचवड महानगरपालिका
मध्यवर्ती भांडार विभाग
कोटेशन नोटीस क्र. ७१/२०२४-२५

पिंपरी चिंचवड महानगरपालिकेच्या पीजीआय वायसीएम रुग्णालयातील रक्तपेढी विभागाकरिता आवश्यक साहित्य/ उपकरण खरेदी करणे कामी ई निविदा प्रस्तावित करावयाची आहे. सदर साहित्य खरेदी करणेकामी इच्छुक पुरवठाधारक यांचेकडून खालील नमूद साहित्याचे दर मागविण्यात येत आहे.

अ.क्र.	साहित्याचे नाव	संख्या	प्रतिनग र.रु सर्व करांसह
१.	Platelet incubator agitator (सोबत जोडलेल्या स्पेसिफिकेशन प्रमाणे)	०१ नग	

कोटेशन नोटीसच्या नियम व अटी खालीलप्रमाणे आहेत.

- पुरवठाधारकाने कोटेशन सादर करताना सर्व कागदपत्रे एका लिफाफ्यात व फक्त दरपत्रक दुस-या लिफाफ्यात दि- २४/०७/२०२४ ते दि- २९/०७/२०२४ अखेर दु. ३.०० वाजेपर्यंत मा. उपआयुक्त (भांडार) यांचे नावे मध्यवर्ती भांडार विभागात सादर करावेत.
- पुरवठा धारकाने आपले दर सर्व करांसह (जीएसटी) देणेत यावे.
- दरपत्रकासोबत शॉप ऍक्ट, जीएसटी नोंदणी प्रमाणपत्र, पॅन कार्ड प्रत तसेच स्पेसिफिकेशनमधील मागणी करण्यात आलेली सर्व प्रमाणपत्रे सादर करणे बंधनकारक आहे.

सही/-
उप आयुक्त (भांडार)
पिंपरी चिंचवड महानगरपालिका,
पिंपरी 411018

जा.क्र.मभां/२२/कावि/ ३४ /२०२४
दि. २४/०७/२०२४

Sr. No.	Technical Specification for Platelet Incubator Agitator
1	Should have an input voltage of 230 +/- 10% V/50-60Hz, Single phase Ac. Should be table top
2	Should have single pane tempered glass door for safety and better visibility of platelet bags
3	Should have built in microcontroller-based temperature recorder and controller unit (TRCU) positioned at eye level for better visibility
4	Must have six inch, 7-day, ink-less pressure sensitive circular chart recorder and accuracy should be +/- 1 degree Celsius
5	Should be able to maintain a temperature of 22 degrees with +/- Degrees variations
6	Should have display of 4*7 Segment LED with 0.5°C display resolution
7	Should have refrigeration with Thermo electric cooling method & 4W LED tube for uniform lighting
8	Should have chart change and set button
9	Should comply IEC safety /EMI/EMC standards, ISO 9001, ISO 13485 certified and CE marked
10	Should have chamber mounted electrical outlet agitator
11	Should have built in voltage stabilizer
12	Should have CFL lamp inside the incubator
13	Should have magnetic gaskets for doors to prevent loss of air
14	Must incorporate a inbuilt battery backup of 2 hours to ensure continuous operation of the TRCU during power failure
15	Should have CFC free PU Insulation
16	Should be able to store minimum of 24/48 platelet bags
17	Should have 4 trays and be able to store minimum 24 platelet bags & trays to store 48 platelet bags
18	Should be flat bed agitator
19	Should have 65 +/-7 cycles motions per min
20	Should have agitation pause function for 10 sec
21	Should have LED indication for blown fuse
22	Should have audible intermittent beeps for power failure
23	Should have SS304 grade sliding stainless steel trays
24	Should comply IEC safety/EMI/EMC standards, ISO 9001, ISO 13485 and CE marked
25	Should have power supply of 220-240 volt at 50Hz single phase
26	Supplied with Servo voltage stabilizer capacity 5 KVA , Single Phase of reputed brand.